



**OFFICE OF THE MANAGING DIRECTOR
ASSAM MEDICAL SERVICES CORPORATION LIMITED
Central Drug Ware House Campus, Patherquerry, Narengi, Guwahati, Assam-781026
Email id: md-amscl@assam.gov.in**

ECF.:299367/I/181982/2023

CORRIGENDUM NO.1

E-TENDER FOR SUPPLY OF VARIOUS EQUIPMENTS

This has reference to the e-tender No: AMSCL/PROC-EQUIP/2023-24/0513/1481 Date: 04/05/2023 invited from manufacturers or authorized distributors / dealers for entering into one financial year rate contract for “**SUPPLY OF VARIOUS EQUIPMENTS**”. Technical specification amendments have been incorporated in the tender as detailed in **Annexure-I**. The details of amendments can be obtained from the website <http://assamtenders.gov.in>.

All other terms & conditions of the tender referred to above shall remain unchanged.

The last date of bid submission is extended to 02/06/2023 up to 2:00 PM. The bids will be opened on the same day at 4:30 PM.

-Sd/-
(Dr. Robin Kumar, IPS)
Managing Director, AMSCL, Assam

Annexure-I

Technical Corrigendum-1

Tender No: AMSCL/PROC-EQUIP/2023-24/0513/1481 dated:04/05/2023 for Supply of Various Equipments.

Note: Amended Technical Specifications are in Bold italic letters.

Name of Item	Sr. No.	AMSCL Technical Specifications	Technical Specifications quoted by Bidder	Deviation (If Any)	If YES: How the quoted specifications is superior or inferior to the quoted specifications
Nebulizer	1	<i>Medicine cup capacity of minimum 5 ml or more</i>			
	2	Weight(lbs,kg)<2 kg			
	3	Should operate on 220-230 VAC,50hZ,5A plug			
	4	Electrical protection by resettable over current breakers or replaceable fuses fitted in both live and neutral lines.			
	5	Provided With necessary accessories: Nebulisation mask (Infants, Paediatric and adult size each 02 Nos. free of cost).PVC tubing for nebulizer(Two pair extra free of cost),Cable cord-01 No.			
	6	USFDA/CE/BIS Certified and Certificate need to be submitted.			
	7	The manufacturer should be ISO 13485 Certified and Certificate need to be submitted.			
	8	Warranty: 01 Years			
Hub Cutter	Sr. No.	Technical Specification			
	1	<i>Should have Stainless Steel Body of SS-304 grade/ABS.</i>			
	3	Should have minimum 400 ml Container			

Name of Item	Sr. No.	AMSCL Technical Specifications	Technical Specifications quoted by Bidder	Deviation (If Any)	If YES: How the quoted specifications is superior or inferior to the quoted specifications
	4	Should have stainless steel alloy adjustable blade(Non-Magnetic)			
	7	Should have Easy removable cap			
	9	Should have plastic holder at the end of the cutter			
	10	Needle Container should have Bio-Hazard symbol of (Black/Red) on the front & Back.			
	11	Manufacturer should be ISO 13485 certified and Certificate need to be submitted.			
	12	Warranty : 01 Year.			
Oxygen Hood	Sr. No.	Technical Specification			
	1	Transparent Polycarbonate unbreakable single molded.			
	2	Silicone rubber Neck port adjustment to minimize the wastage of Oxygen and adjustment to ensures use in Neonate/Infant/Paediatric Patients.			
	3	Oxygen Inlet Port			
	4	Accessories to be provided: Tubings-01 No. free of cost			
	5	The manufacturer should be ISO 13485 certified and Certificate need to be submitted.			
	6	Warranty-01 Years.			
Resuscitator bag-Adult	Sr. No.	Technical Specification			
	1	Resuscitator Bag with valve - Adult(silicone)			
	2	Manufacturer should be ISO 13485 Certified and Certificate need to be submitted.			
	3	<i>Should be EU CE from notified body/USFDA/CE/BIS Certified and Certificate need to be submitted.</i>			
Resuscitation Kit (Adult)	Sr. No.	Technical Specification			
		The Resuscitation Kit (Adult) contains following items:			

Name of Item	Sr. No.	AMSCL Technical Specifications	Technical Specifications quoted by Bidder	Deviation (If Any)	If YES: How the quoted specifications is superior or inferior to the quoted specifications
	1	Adult Resuscitator Bag:Rugged ,100% Autoclavable,Reusable Silicone Bellow with rebreathing Valve & face mask,Foldable,360 degree swiveling patient standard connector,Reservoir bag and 1.5 mtrs PVC Oxygen tubing(1 Unit)(Brain Circuit)			
	2	<i>Silicone Face Mask of all adult sizes i.e Size 1,2 & 3 (1 no. each)</i>			
	3	<i>Adult Laryngoscope Set : 1.Curved Blades of all sizes for adult i.e Size 1,2 ,3 & 4 (1 no. each.) 2.Handle with Cell 3.LED Light Source</i>			
	4	<i>Silicone Airway of all sizes for Adult i.e Size 1,2,3 & 4(1 no. each) & Laryngeal Mask Airway(LMA) Silicone of all sizes for Adult ie.Size 2,3 & 4(1 no. each) - Flexible and disposable (1 No. each sizes)</i>			
	5	<i>Disposable Cuffed E.T tube of all sizes i.e Size 5; 5.5; 6; 6.5; 7; 7.5; 8; 8.5; 9; 9.5 & 10 (1 no. each) for Adult patients.</i>			
	6	Mouth Opener			
	7	All items should be Packed in good quality Rexene Carry Bag			
	8	Manufacturer should be ISO 13485 Certified and Certificate to be submitted.			
	9	Warranty: 1 Year			
Resuscitation Kit (Paediatric)	Sr. No.	Technical Specification			
		The Resuscitation Kit (Paediatric) contains following items:			

Name of Item	Sr. No.	AMSCL Technical Specifications	Technical Specifications quoted by Bidder	Deviation (If Any)	If YES: How the quoted specifications is superior or inferior to the quoted specifications
	1	Paediatric Resuscitator Bag: (of Size 250ml, 500ml, 750ml- 01 no. each) Rugged , 100% Autoclavable,Reusable Silicone Bellow with rebreathing Valve & face mask,Foldable,360 degree swiveling patient standard connector,Reservoir bag and 1.5 mtrs PVC Oxygen tubing(1 Unit)(Brain Circuit)			
	2	Silicone Face Mask of all Paediatric sizes i.e Size 00, 01, 02 & 03(1 no. each)			
	3	Paediatric Laryngoscope Set : 1. Curved Blades of all sizes i.e Size 0 & Size 1(1 no. each) for Paediatric. 2.Handle with Cell 3.LED Light Source			
	4	Silicone Airway of all sizes i.e 2,3,4 & 5 (1 no. each) for Paediatric & Laryngeal Mask Airway(LMA) Silicone of all sizes i.e 1,1.5,2 ,2.5 & 3,(1 no. each) for paediatric -Flexible and disposable (1 No. each sizes).			
	5	Disposable Cuffed E.T tube of all sizes i.e 2.5; 3; 3.5 & 4mm (1 no.each) for paediatric patients.			
	6	Mouth Opener			
	7	All items should be Packed in good quality Rexene Carry Bag			
	8	Manufacturer should be ISO 13485 Certified and Certificate to be submitted.			
	9	Warranty: 1 Year			