





OFFICE OF THE MANAGING DIRECTOR ASSAM MEDICAL SERVICES CORPORATION LIMITED

Central Drug Ware House Campus, Patherquery, Narengi, Guwahati, Assam-781026 Email id: md-amscl@assam.gov.in

ECF No.: 298482/I/179197/2023

CORRIGENDUM NO.2

e-TENDER FOR SUPPLY OF ANESTHESIA WORKSTATION

This has reference to the e-tender No: AMSCL/PROC-EQUIP/2023-24/0467/1421, Dated: 04/05/2023 invited from manufacturers or authorized distributors / dealers for entering into one financial year rate contract for "Supply of Anesthesia Workstation" to health institutions in different districts of Assam."

The following amendment and clarification may be taken note of prior to submission of bids.

All other terms and conditions of the above referred tender shall remain unchanged.

The participating bidder must submit the technical compliance Statement "Annexure XI in the following Format

Technical Specifications

S1 No	Name of the Item	Original Tender Specifications	Now to be read as
1	1.Anaesthesia Workstation-Type A	1.E. Anaesthesia workstation should have inbuilt Ventilator with Colored touch screen 10 inch or more display.	1.E.Anaesthesia workstation should have inbuilt Ventilator with Colored touch screen 08 inch or more display.
2	1.Anaesthesia Workstation-Type A	8.A. Integrated monitor (10" or more color display) for electronic monitoring and display of following set and measured values: Expiratory Tidal Volume, Expiratory Minute volume, PEEP, Peak and Mean and Plateau airway pressure, Frequency, Waveform display for	8.A. Integrated monitor (8" or more color display) for electronic monitoring and display of following set and measured values: Expiratory Tidal Volume, Expiratory Minute volume, PEEP, Peak and Mean and Plateau airway pressure, Frequency,

Airway pressure, flow and	Waveform display for Airway
volume.	pressure, flow and volume.

SD/-(Dr. Robin Kumar, IPS) Managing Director, AMSCL, Assam